

Company Name: ZhongShan Vasaerre Technology Co.,Ltd

Disposable Medical Protective Coverall (Sterilization) Compliance with EN14126 TYPE4 / AAMI Level3 Standard

Main Users Groups:

hospital groups(Switzerland, Italy, Spain,Portugal, Iran & China,etc) private clinics, retail pharmacies, and licensed wholesalers throughout the world

Why Choose Us:

Seamless Sealing Tape Patent Holder

(Sealing Polyester Fibre Patent No. :ZL 2018 2 1410630.6)

Quick Delivery: 1,000,000 PCS in 10 Days.

FOB & CIF Quotation Term

Qualified Certification

- **OBusiness Licence**
- Food and Drug Business License
- Product Introduction
- FDA Certification
- **OSGS** Electrostatic Properties Testing Report
- © CE by OCT (Compliance With Type4/AAMI Level 3)
- OCT Testing Data (Italian Certificate Organization)
- ©GTT Testing Report (China GB 19082-2009 Standard)
 (The Applicant Company Name:Beautiful Nonwoven Company Limited, our Melt-Blown Fabrics Supplier)
- ODeclaration of Conformity Authorization Letter

HQ40" container loading q'ty: 24300pcs totally 972cases, each case 25 pcs/1size in case Delivery time: 1-2days for 300,000 pcs



ZhongShan Vasaerre Technology Co.,Ltd

(OCF Brand Owner, Seamless Sealing Tape Patent Holder & Exporter)

统一社会信用代码

91442000MA4UNJG51Q

名

中山华沙利科技有限公司

类 型 有限责任公司(自然人投资或控股)

法定代表人 王波

经营范围

新型服装功能面料、智能服装、防辐射服装、智能加热服、防弹服、防化服、防护服装用具、野营装备的设计、研究、开发;纺织材料的检验、检测;设计、生产、研发、销售(含网上销售):服装、体育运动器材、第一类医疗器械(医疗器械生产;第二类、第三类医疗器械经营;货物或技术进出口(国家禁止或涉及行政审批的货物和技术进出口除外)。(上述经营范围涉及货物进出口、技术进出口)(依法须经批准负项目,经相关部门批准后方可开展经营活动。)

注册资本 人民币壹仟万元

成立日期 2016年04月14日

营业期限长期

住 所 中山市沙溪镇豪吐龙山路1号3楼

登记机关



经营者名称:中山华沙利科技有限公司

所:中山市沙溪镇豪吐龙山路1号3楼/

经营场所:中山市沙溪镇豪吐龙山路1号3楼

仓库地址:中山市沙溪镇豪吐龙山路1号3楼

法定代表人: 王波 乙乙乙乙乙

负责人:主波。

质量负责人:*****

主体业态: 医疗器械经营企业

社会信用代码: 91442000MA4UNJG51Q

(身份证号码)

投诉举报电话: 12315

中山市市场监督管理局

日常监督管理机构:

日常监督管理人员:由日常监督管理机构指定

有效期至 / 2025 年 4 月 22 日

(注:如需延续有效期,应在有效期届满30日前申请延续)

许可证编号: JY94420121044596

许可范围:

备案内容: 第二类医疗器械批发(长期有效)**

发证机关:

签发人:

2020年

NO. 1094405



OCF Protective Coverall Photographed at a Hospital in Italy



Coverall Material Made of: 67g White Composite Film + 30g Nonwoven Fabric + 35g Film + 32g Glue + 2.0cm EVA 0.12mm Sealing Line



Product Photographs

















Packing Details: 25 Units in 1 Carton

Size	Gross Weight	Carton Dimension
Size	1 Case	W/L/H
175CM(L)	7.9 KGS	65*34*28.5CM
180CM(XL)	8.2 KGS	65*34*28.5CM
185CM(XXL)	8.6 KGS	65*34*28.5CM

^{*}Gross Weight for Each Case: Including the Coverall & Paper Carton.

^{*}HQ40" loading Q'ty: 24300pcs



Unit Vacuum-packing







Carton Photo



OCF Protective Coverall at China Wu Han Hospital
0% Influence Rate





OCF Protective Coverall Photographed at a Hospital in Canadian









SGS Electrostatic Properties Testing Report AAMI Level 3









Test Report SL52025256525301TX Date:May 11,2020 Page 1 o ZHONGSHAN VASAERRE TECHNOLOGY CO.,LTD 3RD FLOOR,NO.1 HAOTULONGSHAN ROAD,SHAXI TOWN,ZHONGSHAN,GUANGDONG,528471 CHINA Page 1 of 2

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A)Medical Protective Clothing

Style No. : One-piece,170cm, 175cm, 180cm, 185cm, 190cm

(A)PE, polyeth yle ne polypro pylene (A)white Composition

Sample Color

: ZHONGSHAN VASAERRE TECHNOLOGY CO.,LTD Manufacturer

Country of Destination

Supplier : ZHONGSHAN VASAERRE TECHNOLOGY CO.,LTD

Proposed Care Instruction:

Test Performed : Selected test(s) as requested by applicant

Sample Receiving Date : May 06, 2020

: May 07, 2020 - May 11, 2020 Testing Period

: Unless otherwise stated the results shown in this test report refer only to the Test Result(s)

sample(s) tested, for further details, please refer to the following page(s).

Personal Protective Equipment - Protective Clothing- Electrostatic Properties- Part 5: Material Performance and Design requirements

Conclusion Remark

Clause 4.2.1 Electrostatic Material M

Requirements

Remark(s): M=Meet General Requirement

Signed for and on behalf of

SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center

Hattie Chen (Account Executive)



No.889, Yehen Roed, Xuhui Dietrici Shanghai, China 200233 中国・上海・徐汇区宣山路889号3号標 邮場: 200233

1 (86-21) 61402666

1 (86-21) 64958763

e ags.china@ags.com

Member of the SGS Group (SGS SA)





Test Report

SL52025256525301TX

Date:May 11,2020

Page 2 of 2

Test Result

Personal Protective Equipment - Protective Clothing- Electrostatic Properties- Part 5: Material Performance and Design requirements

EN 1149-5:2008

Clause 4.2.1 Electrostatic Material Requirements

(EN 1149-1:2006; Test Condition: Applied Voltage and time: 100V for 15 seconds

Temperature: 23±1℃; Relative Humidity: 25±5%)

(A) Unit Face Side Back Side Requirement

As Received

Max. 2.5x10⁹ Ω

Surface Resistance Ω 8.1x10^7 7.6x10^7 on at least one

surface

Conclusion Pass

The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

End of Report



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overlead, available on requests or accessible and the printed of the Company subject to the company of the Company of

| 5 Building No.681 Yorken Road Juhin District Shanghai China 200233 1 (86-21) 61402696 1 (86-21) 64958763 www.sgsgroup.com.cn 中国・上海・徐仁区宜山路889号3号楼 解稿: 200233 1 (86-21) 61402696 1 (86-21) 64958763 e egs.china@egs.com

Member of the SGS Group (SGS SA)



FDA Registration No.: 10065113



Fiscal Year 2020

CERTIFICATION OF REGISTRATION

This certifies that:

ZHONGSHAN VASAERRE TECHNOLOGY CO., LTD.

3rd floor, No.1, Haotulongshan Road, Shaxi town,

Zhongshan, Guangdong, 528471, CHINA

has completed the FDA Establishment Registration and Device Listing with the US Food & Drug Administration, through Shenzhen Huide Medical Device Certification Service Co., Ltd.

Owner/Operator Number: 10065113

Listing Number: Product Code: Device Category Product Name:

D380673 LYU ACCESSORY, SURGICAL Medical Protective
APPAREL Clothing

Shenzhen Huide Medical Device Certification Service Co., Ltd. will confirm that such registration remains effective upon request and presentation of this certificate until end of the calendar year stated above, unless said registration is terminated after issuance of this certificate. Shenzhen Huide Medical Device Certification Service Co., Ltd. makes no other representations or warranties, nor does this certificate make any representation or warranties to any person or entity other than the named certificate holder, for those sole benefit it is issued. This certificate does not denote endorsement or approval of the certificate-holder's device or establishment by the U.S. Food and Drug Administration. Shenzhen Huide Medical Device Certification Service Co., Ltd. assumes no liability to any person or entity in connection with the foregoing.

Pursuant to 21 CFR 807.39, "Registration of a device establishment or assignment of a registration number does not in any way denote approval of the establishment or its products. Any representation that creates an impression of official approval because of registration or possession of a registration number is misleading and constitutes misbranding."

The U.S. Food and Drug Administration does not issue a certificate of registration, nor does the U.S. Food and Drug Administration recognize a certificate of registration, Shenzhen Huide Medical Device Certification Service Co., Ltd. is not affiliated with the U.S. Food and Drug Administration.

Shenzhen Huide Medical Device

Certification Service Co., Ltd.

16C, Building 1, Sunshine Green,

No. 3003 Aiguo Road, Cuizhu Street, Luohu District, Shenzhen

Issue Date: March 27, 2020

Expiration Date: 31 December, 2020



Compliance With Type 4 Standard Testing Report by OCT (Italian Certificate Organization)







Test Report No.:OCT20200326008P

PPE Report EN 14126

Protective clothing -

Performance requirements and tests methods for protective clothing against infective agents

Testing laboratory:	Oct Technology Testing Co., Ltd.
Address:	637,No. 56, zhongyun Road,Panyu District,Guangzhou,Guangdong Province,China
Report body	Oct Technology Testing Co., Ltd.
Address	637,No. 56, zhongyun Road,Panyu District,Guangzhou,Guangdong Province,China
Applicant	ZHONGSHAN WARAOLI TECHNOLOGY CO. LTD.
Address	3rd floor, no.1, haotulongshan road, shaxi town, zhongshan city
Standard	EN 14126:2003/AC:2004
Test Result	Compliance with
	EN 14126:2003/AC:2004
Procedure deviation	N.A.
Trade Mark:	N.A.
Non-standard test method	N.A.
Type of test object:	Disposable Medical Protective coverall
Model/type reference:	one-piece: 175cm、180cm、185cm, split: 175cm、180cm、185cm
Manufacturer:	ZHONGSHAN WARAOLI TECHNOLOGY CO. LTD.
Address:	3rd floor, no.1, haotulongshan road, shaxi town, zhongshan city

Oct Technology Testing Co., Ltd.
637,No. 56, zhongyun Road,Panyu District,Guangzhou,Guangdong Province,China
TEL:020-89015652,888@oucetesting.com
www.oucetesting.com

Page 2 of 10





Test Report No.:OCT20200326008P

General remarks

This report shall not be reproduced except in full without the written approval of the testing laboratory.

The test results presented in this report relate only to the item(s) tested.

"(see appended table)" refers to a table appended to the report.

"(see remark #)" refers to a remark appended to the report.

"(see Annex#)" refers to an annex appended to the report.

Throughout this report a comma (point) is used as the decimal separator.

Remark:

Photos view:

(See appendix 1)

Copy of marking plate:

(See appendix 2)

Oct Technology Testing Co., Ltd.
637,No. 56, zhongyun Road,Panyu District,Guangzhou,Guangdong Province,China
TEL:020-89015652,888@oucetesting.com
www.oucetesting.com

Page 3 of 10





Test Report No.:OCT20200326008P

Possible test case verdicts : test case does not apply to the test object :	
test case does not apply to the test object	
	N (.A.)
test object does meet the requirement:	P(ass)
test object does not meet the requirement :	F(ail)
	7,No. 56, zhongyun Road,Panyu strict,Guangzhou,Guangdong Province,China
Reported by : Signature Approved by :	2020,3.26 Date 2020,3.26

Oct Technology Testing Co., Ltd.
637,No. 56, zhongyun Road,Panyu District,Guangzhou,Guangdong Province,China
TEL:020-89015652,888@oucetesting.com
www.oucetesting.com

Page 4 of 10





Test Report No.:OCT20200326008P

	EN 14126				
Clause	Requirement – Test	Result - Remark	Verdict		
4	Requirements				
4.1	Materials requirements	~.	Р		
4.1.1	General		Р		
	If the care instructions indicate that the clothing car be cleaned and reprocessed at least five times, protective clothing materials shall be submitted to five cleaning and reprocessing cycles according to the manufacturer's care instructions before testing	0	Р		
	If the care instructions specify a lower number of cleaning/reprocessing cycles, then materials shal besubmitted to the number of cleaning/reprocessing cycles indicated.		P		
	Unless otherwise stated in the relevant test procedure, the specimens shall be conditioned for at least 24 h inan atmosphere of (20 ± 2) °C and (6 ± 5) % relative humidity before testing. Tests shall be carried out in thesame atmosphere or within 5 min of removing the sample from the conditioning atmosphere.	55 II	P		
4.1.2	Mechanical and flammability requirements		Р		
	The materials shall be tested and classified in accordance with the test methods and performance classification system specified in the relevant clauses of EN 14325.	ce	P		
4.1.3	Chemical requirements	The state of the s	Р		
4.1.4	Performance requirements against penetration by infective agents		Р		
4.1.4.1	Resistance to penetration by contaminated liquids under hydrostatic pressure	Class!	Р		
	When tested in accordance with ISO/FDIS 16603 and ISO/FDIS 16604 the material shall be classified	3	Р		
	according to the levels of performance given in Table 1, as obtained in the bacteriophage test				

Oct Technology Testing Co., Ltd.
637,No. 56, zhongyun Road,Panyu District,Guangzhou,Guangdong Province,China
TEL:020-89015652,888@oucetesting.com
www.oucetesting.com

Page 5 of 10





Test Report No.:OCT20200326008P

			EN 14126					
Clause	Requiremen	Requirement – Test Result - Remark						
Table 1 –	- Classification o		penetration by contaminate ure (ISO/FDIS 16604)	ed liquids under hydrostatic	-			
		Class	Hydrostatic pressure at w the material passes the tes					
		6	20 kPa					
		5	14 kPa					
		4	7 kPa					
		3	3,5 kPa					
		2	1,75 kPa					
		1	0 kPa a					
	# the		at the material is only expose essure of the liquid in the test					
4.1.4.2		contact with	n by infective agents du substancescontaining	eto Class 1	Р			
		II be classifie	nce with Annex A the ed according to the level	Breakthrough s of time:15min	Р			
Table 2 —	- Classification of	f resistance to		nts due to mechanical contact quids				
		Class	Breakthrough time, t					
		6	t > 75					
		5	60 < t ≤ 75					
		4	45 < t ≤ 60					

	1 ≤ 15 min		
4.1.4.3	Resistance to penetration by contaminated liquid aerosols	Class 1	Р
	When tested in accordance with ISO/DIS 22611 the material shall be classified according to the levels of performance given in Table 3.	Penetration ratio log 2.3	Р

 $30 < t \le 45$

 $15 < t \le 30$

Oct Technology Testing Co., Ltd.
637,No. 56, zhongyun Road,Panyu District,Guangzhou,Guangdong Province,China
TEL:020-89015652,888@oucetesting.com
www.oucetesting.com

3

2

Page 6 of 10





Test Report No.:OCT20200326008P

			EN 14126		
Clause	Requireme	ent – Test		Result - Remark	Verdict
Table 3	— Classificati	on of resistanc	e to penetration by cont	aminated liquid aerosols	
		Class	Penetration ratio (log)		
		3	log > 5		
		2	3 < log ≤ 5		
		1	1 < log ≤ 3		
4.1.4.4	Resistance particles.	to penetration	by contaminated solid	Class 1	Р
			ce with ISO/DIS 22612 sified according to the	Penetration log cfu 2.5	Р
	performand	ce given in Tab	ole 4.		
Table 4	— Classificat	tion of resistan	ice to penetration by cor	ntaminated solid particles	
		Class	Penetration (log cfu)	_	
		3	<1	_	
		2	1 < log cfu ≤ 2		
		1	2 < log cfu ≤ 3		
4.2	Performand		ts for seams, joins and		Р
		ainst infective	plages of protective agents shall fulfil the		Р
		ngth shall be c	clauses of prEN 14325 lassified according to 5.5	5	
4.3	Whole suit	requirements		Type 3	Р
	fulfil the rel whole suitr	evant requirentequirements s	st infective agents shall nents of EN 340 and the pecified in the relevant stective clothing (see	Relevant standard EN 466	Р

Oct Technology Testing Co., Ltd.
637,No. 56, zhongyun Road,Panyu District,Guangzhou,Guangdong Province,China
TEL:020-89015652,888@oucetesting.com
www.oucetesting.com

Page 7 of 10





Test Report No.:OCT20200326008P

EN 14126					
Clause	Requirement - Test	Result - Remark	Verdict		

	Туре	of clothing	Relevant standard		
	type	1a, 1b, 1c, 2	EN 943-1 (EN 943-2 for ET suit	s)	
	type	3	EN 466		
	type	4	EN 465		
	type	5	prEN ISO 13982-1		
	type	6	prEN 13034		
	partia	al body protection	EN 467		
5		Marking			Р
		CANADA CONTRACTOR OF THE PROPERTY OF THE PROPE	be marked in accordance with uirements of the relevant ve clothing.		P
		The marking of pr	otective clothing against infective ain the following additional		Р
		a) the number of t	his European Standard;	EN 14126:2003	Р
			ective clothing, as specified in suffix "-B", e.g. type 3-B;	Type 3	Р
		c) the pictogram "hazard"	protection against biological	⊕	Р
6		Information suppli	ed by the manufacturer		Р
			r the user shall be worded clearly ly and be understandable by a		Р
		against infective a informationrequire	or the user of protective clothing gents shall contain all the d by EN 340 and by the relevant specific type of chemical		Р

Oct Technology Testing Co., Ltd.

637,No. 56, zhongyun Road,Panyu District,Guangzhou,Guangdong Province,China

TEL:020-89015652,888@oucetesting.com

www.oucetesting.com

Page 8 of 10





Test Report No.:OCT20200326008P

EN 14126				
Clause	Requirement – Test	Result - Remark	Verdict	
	protective clothing. In			
	addition it shall contain the following information:			
	a) the number of this European Standard;	EN 14126:2003	Р	
	b) the type designation, e.g. type 3-B;	type 3	Р	
	c) the biological agents against which the protective clothing has been tested. This information shall be expressed as performance levels, as specified in		Р	
	4.1.4.1 to 4.1.4.4 for the relevant types of biological	al		
	challenge;			
	d) all other relevant information on performance levels, preferably as a Table;		P	
	e) the information necessary for trained persons about:		Р	
	application and limitations of use (temperature range, etc.);	10-40℃	Р	
	if relevant, checks to be carried out by the wearer before use;	5	Р	
	fitting and adjustments, and any accessories needed to provide the claimed level of protection;		Р	
	use;		Р	
	maintenance, cleaning and disinfection;		Р	
	storage;		Р	
	if relevant, a warning against problems likely to be encountered;		Р	
	if relevant, illustrations, part numbers and marking of spare parts, etc.	3	Р	
	disposal after use.	Concentrated destruction	Р	

Oct Technology Testing Co., Ltd.
637,No. 56, zhongyun Road,Panyu District,Guangzhou,Guangdong Province,China
TEL:020-89015652,888@oucetesting.com
www.oucetesting.com

Page 9 of 10





Appendix 1

Photo of EUT



OCT TECHNOLOGY TESTING CO., LTD.

Page 10 of 10













TEST REPORT

(Electronic version)



VERIFICATION WEBSITE: www.gttc.net.cn VERIFICATION CODE: EFJJ-4606-24



ISSUE DATE 签发日期:2020-02-07

APPLICANT:
委托单位:Beautiful Nonwoven Company Limited
佛山市南海必得福无纺布有限公司ADDRESS:
地 址:Shatou Jiujiang Town, Nanhai Foshan City, Guangdong
广东省佛山市南海区九江镇沙头石江工业区

客户认定信息 APPLICANT PROVIDED SAMPLE DESCRIPTION: LF20106 67gsm 白色SF透气复合材料 200cm×100cm

颜色 COLOR: 白色 单位质量 WEIGHT: 67gsm

DATE RECEIVED/DATE TEST STARTED: 2020-02-02

样品受理/测试开始日期: 2020-02-02

CONCLUSION 结论:

RESISTANCE TO WATER PENETRATION M 抗渗水性 SURFACE WATER REPELLENCY M 表面抗湿性 BREAKING STRENGTH M 断裂强力 ELONGATION AT BREAK 断裂伸长率 ELECTROSTATIC DECAY M 静电衰减性能 RESISTANCE AGAINST PENETRATION BY SYNTHETIC BLOOD M 抗合成血液穿透性 FILTRATION EFFICIENCY M 过滤效率 WATER-VAPOUR TRANSMISSION RATE M

透湿量

APPROVED BY(签发):

胡万丽 工程师

WanLi Hu ENGINEER

总部:广州市番禺区珠江路1号 花都实验室:广州市花都区狮岭镇旗岭河滨西路1号

电话:020-61994598/61994599 电话:020-37721161













TEST REPORT

(Electronic version)



VERIFICATION WEBSITE: www.gttc.net.cn VERIFICATION CODE: EFJJ-4606-24



ISSUE DATE 签发日期:2020-02-07

NOTE: "M" -MEET THE STANDARD'S REQUIREMENT "F" -FAIL TO MEET THE STANDARD'S REQUIREMENT "--" -NO COMMENT

备注: "M" -符合标准要求 "——" -此项目不判定 "F" -不符合标准要求

REMARK:

REMARK:
ALL THE TESTED ITEMS ARE TESTED UNDER THE STANDARD CONDITION (EXCEPT FOR INDICATION).
本报告中检验检测项目均在相应标准规定的环境条件下进行(有注明的除外)。
COPIES OF THE REPORT ARE VALID ONLY RE-STAMPED.
复印件、副本未重新加盖报告书确认章无效。
THE EXPERIMENT WAS CARRIED OUT AT No. 1, ZHUJIANG ROAD, PANYU DISTRICT, GUANGZHOU, GUANGDONG, P. R. CHINA.
本报告检验检测地址为广州市番禺区珠江路1号。

APPROVED BY(签发): 胡万丽 工程师

WanLi Hu ENGINEER

总部:广州市番禺区珠江路1号

花都实验室: 广州市花都区狮岭镇旗岭河滨西路1号



电话:020-61994598/61994599

电话:020-37721161



3/4











TEST REPORT

(Electronic version)

No:200013643

RESISTANCE TO WATER PENETRATION (kPa) 抗渗水性(kPa)

(GB/T 4744-1997, STEADILY INCREASING WATER PRESSURE 增压法: 6. 0kPa/min, TEMPERATURE OF THE WATER 水温: 20.0℃, FACE DOWN AND FACE SIDE TESTED 正面向下接触水面)

REQUIREMENT 要求 ≥1.67 2# 37.9 (GB 19082-2009) 3# 37.5 4# 34.7 33.9 5# AVERAGE 平均值

SURFACE WATER REPELLENCY

表面抗湿性

(GB/T 4745-1997) (GB/T 4745-1997 EQV等效采用 ISO 4920-1981)

REQUIREMENT 要求 ≥3 GRADE 级 (GB 19082-2009) 4 GRADE 级 4 GRADE 级 2# 4 GRADE 级

BREAKING STRENGTH

断裂强力

(GB/T 3923.1-1997, STRIP METHOD 条样法, THE DISTANCE BETWEEN THE CLAMPS 夹距: 100mm,

RATE 速度:100mm/min)

REQUIREMENT 要求

LENGTH 纵向 128N ≥45N WIDTH 横向 62N ≥45N (GB 19082-2009)

ELONGATION AT BREAK

断裂伸长率

(GB/T 3923.1-1997, STRIP METHOD 条样法, THE DISTANCE BETWEEN THE CLAMPS 夹距:100mm,

RATE 速度:100mm/min)

REQUIREMENT 要求

LENGTH 纵向 72% ≥15% WIDTH 横向 101% ≥15% (GB 19082-2009)

ELECTROSTATIC DECAY

静电衰减性能

(GB 19082-2009 5.10, TEMP 温度: 23.5℃, RH 相对湿度: 50.1%, VOLTAGE APPLIED

施加电压: ±5000V)

REQUIREMENT 要求

≤0.5s MAXIMUM 最大值 0.05s

(GB 19082-2009) RESISTANCE AGAINST PENETRATION BY SYNTHETIC BLOOD 抗合成血液穿透性

(GB 19082-2009 ANNEX A 附录A SURFACE TENSION OF SYNTHETIC BLOOD

合成血液表面张力: 0.042N/m)

REQUIREMENT 要求 6 GRADE 级 ≥2 GRADE 级 (GB 19082-2009)

PAGE 3 UF

总部:广州市番禺区珠江路1号

花都实验室:广州市花都区狮岭镇旗岭河滨西路1号

电话:020-61994598/61994599

电话:020-37721161













TEST REPORT

(Electronic version)

No:200013643

FILTRATION EFFICIENCY

(GB 19082-2009 5.7, AIR FLOW 空气流量:15L/min, AEROSOL 气溶胶颗粒:NaCl, AEROSOL CONCERTRATION 气溶胶浓度:15mg/m³, TEMP 温度:23.5℃RH 相对湿度:38.5%)
REQUIREMENT 要求

MINIMUM 最小值

≥70%

(GB 19082-2009)

WATER-VAPOUR TRANSMISSION RATE

(GB/T 12704-1991, METHOD A 方法A: DESICCANT METHOD 吸湿法, TEST CONDITION 试验箱环境:38.0℃,90.0%)

100%

 $3770g/(m^2 \cdot d)$

REQUIREMENT 要求 ≥2500g/(m² • d) (GB 19082-2009)

Page 4 of

-本报告结束(End of Report)-

总部:广州市番禺区珠江路1号

电话:020-61994598/61994599



Zhongshan vasaerre technology co. LTD

Declaration of Conformity

Zhongshan vasaerre technology co. LTD 3rd floor, no.1, haotulongshan road, shaxi town, zhongshan city We declare that the following product:

Disposable Medical Protective coverall

Models No: one-piece: 175cm, 180cm, 185cm, split: 175cm, 180cm, 185cm

Described above is in conformity with the following directive (s):

Personal Protective Equipment 2016/425

Relevant standard (s):

EN 14126:2003+AC:2004, EN ISO 13982-1:2004+A1:2010

The reference of the File identified with the No: MESTTCF-PPE

And we are aware about the contents and information included in the ModCOM04.06 Regulation that is considered totally accepted.



Date of issue

2020.03.18

Stamp and Signature of authorized personnel of Authori















检验检测报告



防伪查询网址: www.gttc.net.cn 防伪码: EJKT-0692-34

共3页 第1页



		No:200027852	共3页	第1页	IN THE PROPERTY.
委托单位	中山华沙利科技有限 地址:中山市沙溪镇	公司 (豪吐龙山路1号3楼			
客户认定信息	医用一次性防护服	3件			
检验性质	委托检测	样品受理/测试开始日期	2020-03-13	报告签发日期	2020-03-18
判定依据	GB 19082-2009 《医	用一次性防护服技术要求	3)		
综合检验结论					
全验检测	检验	检测项目	判定依据		判定
结果	断裂强力		GB 19082-2009	符	合
	断裂伸长率		GB 19082-2009	符	合
	表面抗湿性		GB 19082-2009		合
	抗渗水性		GB 19082-2009		合
	透湿量		GB 19082-2009		合
	抗合成血液穿透性		GB 19082-2009		合
	过滤效率[关键部位]		GB 19082-2009		合
	过滤效率[接缝处]		GB 19082-2009	符	合
备注	本报告中檢验检測項 复印件、副本未重新 本报告检验检测地址	原目均在相应标准规定的可 所加盖报告书确认着无效。 上为广州市番禺区珠江路1	海条件下进行(有注明的 5	的除外)	AF HERE
		(#		が作り	* 5






















样品图片

No:200027852 共3页 第2页



总部;广州市番禺区珠江路1号 花都实验室;广州市花都区绿岭镇旗岭河滨西路1号

电话:020-61994598/61994599 电话:020-37721161

















检验检测报告附页

No: 200027852 共9页 第3页

		2 1 2 1 1 1	2000	
测试方法	标准值及允差	检验检测结果	判定	备注
GB/T 3923.1-1997 夹距:100mm 速度:100mm/min	纵向 ≥45 横向 ≥45	纵向 145 横向 76	符合	
GB/T 3923.1-1997 夹距:100mm 速度:100mm/min	纵向 ≥15 横向 ≥15	纵向 117 横向 105	符合	
GB/T 4745-1997	≥3	1# 4 2# 4 3# 4	符合	
GB/T 4744-1997 增压法:6.0kPa/min 水温:20.0℃ 正面向下接触水面	≥1.67	1# 19.7 2# 20.0 3# 21.2 4# 20.0 5# 22.0 平均值 20.6	符合	
GB/T 12704-1991 方法A: 吸湿法 试验箱环境: 38.0℃,90.0% 透湿时间: 1h	≥2500	6340	符合	
GB 19082-2009 附录A 合成血液表面张力: 0.042N/m	≥2	6	符合	
GB 19082-2009 5.7 空气流量: 15L/min 气溶胶颗粒: NaCl 气溶胶浓度: 15mg/m ³ 温度:24.1°C 相对湿度:38.2%	≥70	最小值 100	符合	
GB 19082-2009 5.7 空气流量: 15L/min 气溶胶颗粒: NaCl 气溶胶浓度: 15mg/m³ 温度:24.3℃ 相对湿度:38.2%	≥70	最小值 99	符合	
20	湯の川の水		W 产品质量	
	S N W	国家游戏员	*	學出
		检	並检測专用	草
	.,,			
	——本报告结束—			
	GB/T 3923. 1-1997 夹距:100mm/min GB/T 3923. 1-1997 夹距:100mm/min GB/T 3923. 1-1997 夹距:100mm 速度:100mm/min GB/T 4745-1997 GB/T 4744-1997 增压法:6. 0kPa/min 水温:20. 0℃ 正面向下接触水面 GB/T 12704-1991 方法温: 38. 0℃, 90. 0% 透湿时间: 1h GB 19082-2009 附录A 合成血液表面张力: 0. 042N/m GB 19082-2009 5. 7 空气流散粒度: 15L/min 气溶胶浓度: 15mg/m³ 温度:24. 1℃ 相对湿度:38. 2% GB 19082-2009 5. 7 空气溶胶浓度: 15mg/m³ 温度:24. 1℃ 相对湿度:38. 2% GB 19082-2009 5. 7 空气溶胶浓度: 15mg/m³ 温度:24. 3℃ 相对湿度:38. 2%	GB/T 3923. 1-1997 夾距:100mm 速度:100mm/min GB/T 3923. 1-1997 大型:100mm 対域度:100mm/min GB/T 4745-1997 場压法:6. 0kPa/min 水温:20. 0°C 正面向下接触水面 ≥1. 67 GB/T 12704-1991 方法A: 吸湿法 试验箱环境: 38. 0°C, 90. 0% 透湿时间: 1h GB 19082-2009 附录A 合成血液表面张力: 0. 042N/m GB 19082-2009 5. 7 空气流量: 15L/min 气溶胶软度: 15mg/m³ 温度:24. 1°C 相对湿度:38. 2% GB 19082-2009 5. 7 空气流量: 15L/min 气溶胶软度: 15mg/m³ 温度:24. 3°C 相对湿度:38. 2%	(BB/T 3923. 1-1997	選出方法

Sealing Polyester Fibre Patent No. :ZL 2018 2 1410630.6)



证书号第8937294号





实用新型专利证书

实用新型名称:一种衣服接缝高强度胶条

发 明 人: 王贵;王志勇;相里海霞

专 利 号: ZL 2018 2 1410630.6

专利申请日: 2018年08月30日

专 利 权 人: 中山华沙利科技有限公司

地 址: 528471 广东省中山市沙溪镇隆盛路巨邦产业园第一幢 7 楼


授权公告日: 2019年06月07日

授权公告号: CN 208949173 U

国家知识产权局依照中华人民共和国专利法经过初步审查,决定授予专利权、颁发实用 新型专利证书并在专利登记簿上予以登记。专利权自授权公告之日起生效。专利权期限为十 年,自申请日起算。

专利证书记载专利权登记时的法律状况。专利权的转移、质押、无效、终止、恢复和专利权人的姓名或名称、国籍、地址变更等事项记载在专利登记簿上。

局长申长雨





第1页(共2页)





中山市突发公共卫生事件一级响应期间 应急医疗器械生产备案凭证

备案编号:粤中市监应急械生产备 2020005 号

中山华沙利科技有限公司			
中山市沙溪镇豪吐龙山路1号3楼			
中山市沙溪镇豪吐龙山路1号3楼			
王波	企业负责人 王贵		
14-	14-02: 医用一次性防护	3服	
产品名称	产品备案/注册号	登载日期	备注
医用一次性防护服	粤中市监应急械备 2020005 号	2020 年 2 月 15 日	
以下空白			
ال علم علم ال			
	中山市 王波 14- 产品名称 医用一次性防护服 以下空白	中山市沙溪镇豪吐龙山路 1 号 在业负责人	中山市沙溪镇豪吐龙山路 1 号 3 楼 中山市沙溪镇豪吐龙山路 1 号 3 楼 王波 企业负责人 王贵 14-14-02: 医用一次性防护服 产品名称 产品备案/注册号 登载日期 医用一次性防护服 粤中市监应急械备 2020 年 2 月 15 日

备案部门(公養和加出 备案日期: 2020年2月第5日 行政审批专用章



中山市突发公共卫生事件一级响应期间 应急医疗器械备案信息表

备案号: 粤中市监应急械备 2020005 号

备案人名称	中山华沙利科技有限公司	
各案人社会信 用代码	91442000MA4UNJG51Q	
备案人注册地 址	中山市沙溪镇豪吐龙山路1号3楼	
生产地址	中山市沙溪镇豪吐龙山路1号3楼	
产品名称	医用一次性防护服	
分类编码	14-14-02	
型号/规格	型号: 连身式、分身式 规格: 160、165、170、175、180、185。	
产品描述	由一种或多种对病毒气溶胶、含病毒液体等具有隔离作用的面料加. 而成的衣服。脱下时,防护衣的外表面不与人体接触。为非灭菌产品一次性使用。	
预期用途	用于医疗机构医护人员穿的职业防护衣。阻止来自患者的病毒随空气或液体向医务人员传播。	
适用标准	GB 19082-2009 医用一次性防护服技术要求	
备案单位 和日期	全案日期: 2020年2月15日	
变更情况	(1) 中 加 专 用 章	
备注	本备案仅在防控新冠肺炎疫情期间有效	

3/7

中山市突发公共卫生事件一级响应期间 应急医疗器械备案凭证

中山华沙利科技有限公司:

在重大突发公共卫生事件一级响应期间,我局根据《广东省药品监督管理局办公室关于重大突发公共卫生事件一级响应期间对医用口罩等防控急需用器械实施特殊管理的通知》(粤药监办许(2020)39号)及其补充通知(粤药监办许(2020)42号)、《广东省药品监督管理局办公室关于明确疫情防控所需第二类医疗器械应急审批备案有关事宜的通知》(粤药监办许[2020]65号)对你单位申请的疫情应急医疗器械:医用一次性防护服予以备案,备案号:粤中市监应急械备2020005号。

本备案仅在防控新冠肺炎疫情期间有效。

