HERA V (HERACARE)

Multi RF System

Efficacy and Effect

- The effect of pain relief **(Approved by KFDA With HERA V)**
  - · RF Vacuum hand-piece (Three types of handpiece : small, middle, large)
- Vaginal tightening
  - · RF Bar hand-piece (The range of operating temperature : 35°C~47°C)
- The effect of muscle pain relief **(Approved by KFDA With HERA V)**
  - · TENS (Transcutaneous Electrical Nerve Stimulation)
  - · Normal mode, Burst mode, Increase mode (Each mode has 1~10 step)

ElimTek

COPYRIGHT © Elimtek. ALL RIGHTS RESERVED

HERA V (HERACARE)

Multi RF System

RF Vacuum

RF Bar

TENS

Efficacy and Effect

RF Vacuum

- RF power can be controlled with 1~10 RF levels
- RF Vacuum can be controlled with four modes and 5 vacuum levels

RF Bar

- Setting the tempreture of hand-piece
- Setting the treatment times
- Display impedance value
- Display the remaining operating time

TENS

- Three modes and 10 steps(from 1 to 10) are available

Safety

- Interlock : RF Vacuum or RF Bar or TENS don't work at the same time.
- Alarm works when operating time end.
- Fuse is used for the prevention of overcurrent.
- Power will be disconnected if RF hand-piece is wrong connected.

Category			Content
Electrical Rating			10~240V, 50/60Hz
50/60Hz Body Size			380 x 384 x 173 mm
Body Weight			5kg
Hand-Piece Size	RF Vacuum	Large	0 85.99 x 205.5 mm
		Middle	0 59.30 x 205.5 mm
		Small	⊠ 32.91 x 205.5 mm
	RF Bar	S-type	20 x 355.4 mm
		I-type	20 x 355.5 mm

ElimTek Elimtek Co., Ltd.

R1403, Woolim Lions Valley 5(o)-cha, 302, Galmachi-ro, Jungwon-gu, Seongnam-si Gyeonggi-do, Republic of Korea

T. +82-31-742-7128 **F.** +82-31-744-2627 **E.** info@elimtek.co.kr **www.elimtek.co.kr**